

Transforming Community Equipment Programme

Product Specification

<u>PRODUCT CATEGORY</u>	Bathing Aids	<u>SPECIFICATION NUMBER</u>	BA09
<u>PRODUCT NAME</u>	Bath Seat – Standard 12 Inch Type	<u>PRODUCT CODE</u>	

FUNCTION:

- Fitted to the non tap end of the bath to facilitate getting in and out of the bath. It provides an intermediary seat within the bath for those wishing to sit near to or in the water.

MATERIAL:

- Slatted, perforated or solid plastic seat with suction padded feet.
- Drainage holes, slots or perforations must not pose a danger of entrapment of body parts
- Where appropriate all materials used must be recyclable

DIMENSIONS:

- Seat Depth minimum 203mm (preferably 280mm)
- Seat Width approx 400mm
- Seat Height 300mm
- Slats to be a maximum of 8mm apart. Holes or perforations to be a maximum of 8mm diameter
- Minimum SWL 140kgs (22 Stones)

FITTING / FIXTURE REQUIREMENTS:

- Fitted by resting on the rims of the bath and securing by brackets designed to wedge the bath board between the rims.
- Clear fitting and safe use instructions to be included
- Recommended cleaning instructions to be included
- Spare fixing brackets to be stocked
- Must comply with Medical Devices Regulations 2002 if applicable
- Where appropriate must be labelled in accordance with CE marking regulations

SERVICING & MAINTENANCE REQUIREMENTS:

- Not applicable